

June 16, 2010

“MONSEL’S BULLETIN”

Dear Valued Customer,

It has come to Premier Medical’s attention that a very small amount of the aluminum caps used to seal the Monsel’s vials are embossed with the words **“FOR ORAL USE”** under the pull tab. Based on the information available at this time, we are estimating that 0.1% of the Monsel’s vials in your inventory and or your customer’s may be labeled incorrectly. We do not believe this presents risk to public safety as this product is used only by trained professionals in the gynecology and dermatology fields. The Monsel’s solution is the same effective formulation manufactured by Premier and will perform as intended during use. However, Premier Medical wanted you and your customers to be aware of the mislabeled caps to eliminate any confusion and prevent any chance of misuse.

The following lots of Monsel’s (item # 9045055) may have the mislabeled caps that were shipped to you between February 23, 2010 and May 30, 2010: **LOTS: M0161, M0163, M0165**

Premier Medical Products Company will replace any vials found embossed with the wording **“FOR ORAL USE”** under the pull tab if you decide to return them. We appreciate your cooperation and support in promptly disseminating this important information to your customers so we can address any inquiries you may receive.

Please contact Toni Banet, Market Services Administrator at 888-773-6872 with any questions or concerns.

Sincerely Yours,



Vince D’Alessandro
Regulatory Manager
Premier Medical Products Company